

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 11/09/2007
Monitor Time: 1/7/2005 2:2

Serial Number: 51446004
CF: 2.87

Ranch home



Inspector Information

Craig Lester
Lester Insp Services
8243 SW 53rd St
Topeka
Kansas-66610
Phone Number: 785-478-2508
License Number: N/A

Customer Information

Ranch home
usa
usa
usa-usa

Basement location



Site & Condition


Wind: NE 10-15mph
Year Built: 1966
SqFt: 1200

Atmospheric Condition: Clear and Dry
Structure Type: Ranch
Monitor Location: Basement

Test Summary

Start Time: 01/01/2005 01:32
End Time: 01/03/2005 01:32
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs

Overall Avg: 2.9 pCi/l
EPA Avg: 3.0 pCi/l



***** 01/01/2005 *****

Time	Counts pCi/l	Flags
02:32	1.7	
03:32	2.1	
04:32	2.1	
05:32	2.1	
06:32	2.8	
07:32	3.8	
08:32	2.1	
09:32	3.1	
10:32	1.4	
11:32	1.4	
12:32	3.1	
13:32	3.1	
14:32	1.0	
15:32	2.1	
16:32	3.1	
17:32	2.8	
18:32	4.2	
19:32	4.2	
20:32	3.5	
21:32	4.2	
22:32	3.5	
23:32	2.1	

***** 01/02/2005 *****

Time	Counts pCi/l	Flags
00:32	4.5	
01:32	2.1	
02:32	4.5	
03:32	3.5	
04:32	1.7	
05:32	2.4	
06:32	1.7	
07:32	3.8	
08:32	3.8	
09:32	3.5	
10:32	4.5	
11:32	3.8	
12:32	3.1	
13:32	2.8	
14:32	1.0	
15:32	1.7	
16:32	3.1	
17:32	3.8	
18:32	2.4	
19:32	3.5	

***** 01/02/2005 *****

Time	Counts pCi/l	Flags
20:32	1.7	
21:32	3.1	
22:32	3.5	
23:32	3.8	

***** 01/03/2005 *****

Time	Counts pCi/l	Flags
00:32	3.5	
01:32	2.8	

Error Flags:
M Motion:

Inspector Signature _____

PC Software Version: 2.0.0.0
Embedded Software Version: 105